



UM-DAP-040



Instructions for Use

UMIC Daptomycin

UMIC Daptomycin for manual susceptibility testing of *Staphylococcus* spp. and *Enterococcus* spp. against the critically important antibiotic daptomycin

For *in vitro* diagnostic use



Language: en

Document history

Title:	Instructions for Use UMIC Daptomycin
Revision:	Revision A (December 2021)
First revision:	December 2021

The following table describes important changes from the previous revision of this document.

Section	Changes
–	No changes: first revision

Table of contents

Document history	2
Table of contents	3
1 Intended use	4
2 Product description / materials	5
3 Media composition	6
4 Shelf life / storage / disposal	7
5 Precautions	8
6 Test procedure	9
6.1 Preparing the samples	9
6.2 Preparing the inoculum	9
6.3 Inoculating the strips	9
6.4 Sealing and incubation	9
6.5 Reading	10
6.6 Analysis	10
7 Evaluating results	11
7.1 MIC value	11
7.2 Skipped wells	11
8 Technical notes	12
9 Quality control	13
10 Clinical and analytical performance	14
11 Warranty	15
12 Limitation	16
13 Symbols	17
14 References	18
15 UMIC Daptomycin short instruction	19
16 Manufacturer	20

1 Intended use

The UMIC Daptomycin is an *in vitro* diagnostic medical device for the quantitative susceptibility testing of clinically relevant fast-growing aerobic Gram-positive bacteria (*Enterococcus* spp. and *Staphylococcus* spp.) against daptomycin using Mueller Hinton Broth, cation adjusted (CAMHB). Susceptibility is detected by determining minimum inhibitory concentration according to EUCAST or CLSI guidelines. Only pure cultures obtained from human test material can be used.

The test is not automated. The device is intended for laboratory professional use only. The results of the test are intended solely as an aid to diagnosis for targeted antibacterial therapy and must not be used as single source for diagnosis, treatment or patient management decision.

2 Product description / materials

Contents

Each pack is sufficient for 40 susceptibility tests and contains:

- 5 UMIC Daptomycin plates, each containing 8 test strips with 12 wells
- Lid and frame
- Instructions for Use
- Certificate of Analysis
- Results Sheet

Additionally required reagents and materials

- Mueller Hinton II (UM-MH-020); manufacturer: Bruker Daltonics GmbH & Co. KG
- NaCl 0.9%
- Humidity chamber, e.g. UMIC Box Daptomycin UM-BOX DAP
- Adjustable pipette, e.g. 20-200 µL, incl. pipette tips

Additional lab materials

- McFarland standard 0.5 or densitometer
- Blood agar or tryptic soy agar (TSA) plates without additives

Note *Use of other agar media requires validation by the user.*

- Incubator
- Inoculation loops
- Marker pen

3 Media composition

Media	Ingredients
NaCl 0.9%	Sodium chloride
Mueller Hinton II (UM-MH-020)	Beef extract Acid hydrolysate of casein Starch

Note *Determination of valid MIC values using the UMIC Daptomycin test system requires the usage of the Mueller Hinton II manufactured by Bruker Daltonics GmbH & Co. KG only. Please contact your local distributor for catalog numbers of reagents and media.*

4 Shelf life / storage / disposal

UMIC Daptomycin test plates remain usable until the indicated expiry date when stored in their original packaging at 15-25 °C. Unused UMIC Daptomycin strips remain usable for up to 2 months after opening when stored in the resealable original packaging.

Store Mueller Hinton II at 2-25 °C. See product label for shelf life of the Mueller Hinton II.

Dispose of transport and secondary packaging of the UMIC Daptomycin according to the local waste disposal regulations.

Follow the instructions in section 5 to dispose of used UMIC test strips.

5 Precautions

- For *in vitro* diagnostic use only.
- Do not pipette reagents by mouth.
- Do not use for other purposes than the intended use.
- Samples, bacterial cultures, and inoculated UMIC Daptomycin test plates / test strips should be treated as potentially biohazardous and should be handled properly by appropriately qualified and trained staff while observing all relevant precautions. Use aseptic technique throughout the whole test procedure. For more information, see the current version of “BioSafety in Microbiological and Biomedical Laboratories, HHS publication (CDC)” or the relevant national regulations.
- If the primary packaging (aluminum composite foil pouch) is damaged, check the integrity of the sealing foil on the UMIC Daptomycin test plates / test strips. If the sealing foil is undamaged, the UMIC Daptomycin test plates / test strips can be used. If the sealing foil is damaged, discard the UMIC Daptomycin test plates / test strips.
- After reading and evaluation of the test result, all samples and inoculated or contaminated items (pipette tips, UMIC Daptomycin test strips) should be autoclaved, incinerated, or treated with a bactericidal disinfectant solution before disposal.
- The Instructions for Use must be strictly adhered to; any deviation can affect the quality of the results.
- The test results should be interpreted by trained staff with experience in microbiology. Clinical background, sample origin, colony and microscopic morphology, serology, and the identification result need to be considered when interpreting the results.

6 Test procedure

This section describes all the steps that are necessary for using the UMIC strips.

6.1 Preparing the samples

Provide a tube with 2-5 mL NaCl 0.9%, pH 5.5-6.5.

Provide a vial of Mueller Hinton II, see section 8.

6.2 Preparing the inoculum

1. Pick several single colonies from an 18-24-hour-old pure culture from blood agar or TSA without additives.
2. Homogenize the colonies well in 2-5 mL NaCl 0.9% until the turbidity matches to a McFarland standard of 0.5.
3. Gram-positive bacteria: Pipette 50 μ L of the bacterial suspension into 5 mL Mueller Hinton II and homogenize well.

6.3 Inoculating the strips

Remove the UMIC Daptomycin test strips from their resealable packaging max. 30 minutes before use.

1. Cut the sealing foil along the strips to be used and remove the strips from the frame.
2. Put the remaining strips back in the resealable packaging immediately and seal the packaging tightly.
3. Insert the strips into the empty frame that is supplied with the pack. The printed marking "DAP" on each strip must be on the left side of the frame.
4. Remove the sealing foil and note the sample number on the respective strip to prevent mixing up of samples.
5. Inoculate each well of each test strip with 100 μ L of the prepared suspension.

Note *After inoculation, check the correct filling of all the wells and take care that no drops of bacterial suspension remain at the edges of the wells. Drops at the well edges may not have contact with the antibiotic. This can result in growth in the respective wells (skipped wells), see section 7.2.*

6.4 Sealing and incubation

1. After inoculation, seal the strips by placing the lid on the frame.
2. Incubate the test strips at 35 ± 1 °C under aerobic conditions for 18 ± 2 hours.

Note *To avoid drying of the media in the wells, which is a risk existing in incubators with ventilation system, incubate the system in a humidity chamber, e.g. using the UMIC Box Daptomycin UM-BOX DAP.*

6.5 Reading

1. Remove the lid.
2. Wipe the bottom of the strips.
3. Visually read the strips with the printed marking "DAP" positioned on the left.

Note *Turbidity = growth/positive. No turbidity = no growth/negative.*

Note *Growth must be visible in the growth control, otherwise the test must be repeated.*

6.6 Analysis

Document the results on the product-specific Results Sheet and interpret the MIC values according to the currently valid international CLSI or EUCAST standards. The Results Sheet is not part of this Instructions for Use.

7 Evaluating results

7.1 MIC value

The minimum inhibitory concentration (MIC) is defined as the lowest concentration of an antibiotic that inhibits visible bacterial growth. Bacterial growth can be observed as turbidity within the medium and/or as a deposit of cells at the bottom of the affected wells of the test strips. MIC determination requires a positive (turbid) growth control. MIC determination is based on the broth microdilution procedure which allows for a quantitative test result regarding susceptibility against daptomycin for the tested isolate.

The MIC obtained should be interpreted according to the current EUCAST or CLSI interpretive criteria.

7.2 Skipped wells

Occasionally, skipped wells can occur within a test strip. A skipped well is characterized by lack of growth in one or more wells of an antibiotic serial dilution while the wells of the next lower/higher concentration show growth. The reasons for skipped wells are diverse: heteroresistance, inhomogeneous inoculum or inhomogeneous inoculation of the wells, etc.

For UMIC Daptomycin test strips, a single skipped well in a serial dilution can be ignored. The concentration of the well after which no more visible growth is detectable can be read as the MIC value. If several skipped wells occur in a serial dilution, the sample and test preparation must be repeated.

8 Technical notes

For best results, carefully take into account the following remarks in the Instructions for Use.

- UMIC Daptomycin test strips are for single use only. Do not re-use!
- Use pure cultures from blood agar or TSA without additives that are not older than 24 hours.
- After inoculation, check the correct filling of all the wells and take care that no drops of bacterial suspension remain at the edges of the wells. Drops at the well edges may not have contact with the antibiotic. This can result in growth in the respective wells (skipped wells), see section 7.2.
- Use NaCl 0.9%, pH 5.5-6.5.
- Use only the Mueller Hinton II manufactured by Bruker Daltonics GmbH & Co. KG.
- Pre-heat the Mueller Hinton II before use. Incubate at 35 ± 1 °C for 1 hour.
- Make sure to adjust the NaCl suspension exactly to McFarland standard 0.5. Homogenize the prepared suspension sufficiently.
- Inoculate a blood agar plate or TSA without additives with the bacterial suspension for purity control.
- Do not damage the sealing foil of the adjacent strips. Strips with damaged sealing foil should not be used later because there could be a loss of drug activity.
- Ensure that the test strips are aligned correctly when inserting them into the frame. The printed marking "DAP" on each strip must be on the left side of the frame.
- Follow the specified incubation times. The incubation time must not be less than 16 hours for *Staphylococci* spp. and 24 hours for *Enterococci* spp.
- To avoid drying of the media in the wells, incubate the system in a humidity chamber, e.g. UMIC Box Daptomycin UM-BOX DAP.

9 Quality control

Strips and reagents are subjected to several quality inspections, systematically executed during various production stages. Bacteriological quality control can be performed using the following strains:

Strains	ATCC¹ No.	DSMZ² No.	NCTC³ No.
<i>Staphylococcus aureus</i>	ATCC 29213	DSM 2569	NCTC 12973
<i>Enterococcus faecalis</i>	ATCC 29212	DSM 2570	NCTC 12697

In the evaluation of the control strains, the MIC values of the control strains used should fall within the quality control ranges of the respective drug tested as specified in the relevant standard.

¹ATCC=American Type Culture Collection

²DSMZ=Deutsche Sammlung von Mikroorganismen und Zellkulturen GmbH (German Collection of Microorganisms and Cell Cultures Ltd.)

³NCTC=National Collection of Type Cultures

10 Clinical and analytical performance

The general requirements of ISO 20776-1 ff. and/or EUCAST or CLSI apply.

According to the specifications of the standard EN ISO 20776-2, the performance of a product for the antibacterial susceptibility testing is given if, on the one hand, at least 95% of the MIC values of an antibiotic relevant for a test strain lie within the corresponding quality control range or reflect the corresponding resistance phenotype and, on the other hand, the MIC values of an antibiotic deviate from the calculated MIC modal value (or MIC modal range) by a maximum of +/- one dilution level.

11 Warranty











The performance data of the UMIC Daptomycin test strips were determined on the basis of these Instructions for Use. Deviations from and changes to the test procedure may impact the quality of the results. Any claims for damages are excluded in such cases.

We would like to point out that all serious incidents related to this product must be reported to the manufacturer and the competent authority of the EU member state in which the user is located.

12 Limitation

To determine valid MIC values using the UMIC Daptomycin test system, Mueller Hinton II manufactured by Bruker Daltonics GmbH & Co. KG must be used.

13 Symbols


Symbol	Description
	Do not re-use
	Contains sufficient for <n> tests
	Temperature limit
	Consult Instructions for Use
	Caution
	Use-by date
CE	CE marking according to IVDD 98/79/EC
	Batch code
	<i>In vitro</i> diagnostic medical device
	Catalog number
	Manufacturer

14 References

Note *The standards and guidance documents listed below are subject to update.*

- ISO 20776-1 — Clinical laboratory testing and in vitro diagnostic test systems — Susceptibility testing of infectious agents and evaluation of performance of antimicrobial susceptibility test devices — Part 1: Reference method for testing the in vitro activity of antimicrobial agents against rapidly growing aerobic bacteria involved in infectious diseases.
- European Committee on Antimicrobial Susceptibility Testing (EUCAST), Breakpoint tables for interpretation of MICs and zone diameters.
- Methods for Dilution Antimicrobial Susceptibility Tests for Bacteria That Grow Aerobically; CLSI Document M07
- Performance Standards for Antimicrobial Susceptibility Testing; CLSI Document M100.

15 UMIC Daptomycin short instruction



16 Manufacturer

**Bruker Daltonics GmbH & Co. KG**

Fahrenheitstraße 4

28359 Bremen

Germany

Phone: +49 421 2205-0

E-mail: micronaut.support@bruker.comWeb: www.bruker.com/microbiology**For Australian customers:****Bruker Pty. Ltd.**

1/28a Albert Street

Preston

Victoria 3072

Australia

Support (hardware and software)Email: biotyper.anz@bruker.com

Phone: +61 (1800) 171-247

SalesEmail: sales.anz@bruker.com

Phone: +61 (1800) 278-537

Fax: +61 (03) 9474-7070

(Based on *Biocentric Instructions for Use UMIC, Revision 02, April 2020*)

Descriptions and specifications supersede all previous information.

© Copyright 2021 Bruker Daltonics GmbH & Co. KG